

Questions with asterisk (\*) are non-essential, i.e, a candidate would not fail if this task is not completed

## LEVEL 3 MACHINE CHECK PRACTICAL – LAMTEC NATIONAL ASSESSMENT – ANAESTHETIC TECHNOLOGY I

### Marking Guide.

Trainee's name.....

Training Hospital..... Date.....

The following check must be carried out between cases:

Circuit	Yes	No
Filter In situ	<input type="checkbox"/>	<input type="checkbox"/>
Leak Test	<input type="checkbox"/>	<input type="checkbox"/>
Unidirectional valves	<input type="checkbox"/>	<input type="checkbox"/>
APL	<input type="checkbox"/>	<input type="checkbox"/>
CO2 Absorber	<input type="checkbox"/>	<input type="checkbox"/>
* Laerdal bag	<input type="checkbox"/>	<input type="checkbox"/>

Clean Face mask	<input type="checkbox"/>	<input type="checkbox"/>
-----------------	--------------------------	--------------------------

Suction	Yes	No
- Adequate length tubing	<input type="checkbox"/>	<input type="checkbox"/>
- Correctly assembled	<input type="checkbox"/>	<input type="checkbox"/>
- Reaches maximum negative pressure = - 60Kpa	<input type="checkbox"/>	<input type="checkbox"/>

MONITORING	Yes	No
Gas sampling connected to filter	<input type="checkbox"/>	<input type="checkbox"/>
SpO <sub>2</sub>	<input type="checkbox"/>	<input type="checkbox"/>
ECG	<input type="checkbox"/>	<input type="checkbox"/>
NIBP	<input type="checkbox"/>	<input type="checkbox"/>
* Temperature	<input type="checkbox"/>	<input type="checkbox"/>
* Nerve Stimulator	<input type="checkbox"/>	<input type="checkbox"/>
* Cables - pressure	<input type="checkbox"/>	<input type="checkbox"/>
* Stethoscope	<input type="checkbox"/>	<input type="checkbox"/>

Questions with asterisk (\*) are non-essential, i.e, a candidate would not fail if this task is not completed

**LEVEL 3 MACHINE CHECK PRACTICAL – LAMTEC  
NATIONAL ASSESSMENT – ANAESTHETIC TECHNOLOGY I**

<b>Airway</b>	<b>Yes</b>	<b>No</b>
Guedal airways - all sizes	<input type="checkbox"/>	<input type="checkbox"/>
E T Tubes	<input type="checkbox"/>	<input type="checkbox"/>
LMA's	<input type="checkbox"/>	<input type="checkbox"/>
1 medium laryngoscope	} tested	<input type="checkbox"/>
1 large laryngoscope		
Syringe	<input type="checkbox"/>	<input type="checkbox"/>
Stylets / bougies	<input type="checkbox"/>	<input type="checkbox"/>
Tape / Ties	<input type="checkbox"/>	<input type="checkbox"/>
* Magills	<input type="checkbox"/>	<input type="checkbox"/>
	<b>Yes</b>	<b>No</b>
<b>IV Cannulation &amp; administration equipment</b>	<input type="checkbox"/>	<input type="checkbox"/>
<b>Suxamethonium and emergency drugs replenished as required</b>	<input type="checkbox"/>	<input type="checkbox"/>

<b>Additional Equipment</b>	<b>Yes</b>	<b>No</b>
* Infusion pump	<input type="checkbox"/>	<input type="checkbox"/>
* Fluid warming devices (if required)	<input type="checkbox"/>	<input type="checkbox"/>
* Patient warming devices (if required)	<input type="checkbox"/>	<input type="checkbox"/>

**If a vaporiser has been changed between cases, the trainee would be required to complete the following extra tests:**

<b>Vaporisers</b>	<b>Replaced vaporiser</b>		
	<b>Yes</b>	<b>No</b>	
Contents	<input type="checkbox"/>	<input type="checkbox"/>	
Filling port	<input type="checkbox"/>	<input type="checkbox"/>	
Maximum setting	<input type="checkbox"/>	<input type="checkbox"/>	

